



安徽江南医疗器械股份有限公司

Anhui JN Medical Device Co.,Ltd.

Address:Caicun Town Jing County Xuancheng City, Anhui, China

Tel: +86 13905632618 +86 563 5456998 Email:xcjn001@163. com

Web: www. jnmedical. com www. syringe. cc

CERTIFICATE OF ANALYSIS

Product Name	1ML Disposable Syringe	Quantity:	500000PCS
Batch Number	20210101	Manuf. Date:	2021-01
Test Basis	ISO7886-1-2017	Expiry Date:	2025-12
Test Items	Requirements	Results	
Capability is more or less	Should be	Pass muster	
Exterior Conic Tie-in	Should be	Pass muster	
Instrumental Thick Capacity	Should be	Pass muster	
Transparence Without Burring Cleanliness Without Burring	Should be	Pass muster	
Lubricant	Should be	Lubricant Collecting	
Leak	Should be	Pass muster	
The Mini-Scale Staff Gauge	Should be	Pass muster	
Additonal Staff Gauge	Should be	/	
Scale Line Distributing Equality	Should be	Pass muster	
The Longest Scale Line Tidily	Should be	Pass muster	
The Shortest Scale Line Be the Half of The Longest Scale Line	Should be	Pass muster	
Scale's Mini-Staff Gauge	Should be	Pass muster	
Scale Value Position	Should be	Pass muster	
Nominal Capacity's total length	Should be	Pass muster	
Fiducial Line and Scale Zero-line	Should be	Pass muster	
Maximum Usable Capacity>10% Than Nominal Capacity	Should be	Pass muster	
Handle	Should be	Pass muster	
Centric Cone-Shaped Position	Should be	Pass muster	
Fastigate Hole Diameter	Should be	≥1.2mm	
Leftover	Should be	Pass muster	
Reducing(oxidizable) matter	Solution used [c(kmnO4)-0.002mol/1] Shall not exceed 0.5ml	0.15ml	
Metalions	The extract shall not contain in total More than 5 ug/ml	Pass muster	
Titration acidity or alkalinty	Shall not more than 1	0.18	
Sterility	Shall be	Pass muster	
Toxicity	Shall be non-toxic	Pass muster	
Pyrogenicity	Shall be non-pyrogenic	Pass muster	
Haemolysis	Percentage of hacmolysis less than 5%	Pass muster	
Conclusion: Measure up			

